

DCI PRODUCT DATA SHEET

CONFIDENTIAL INFORMATION: Information contained in this document is proprietary to DCI. Unauthorized duplication or distribution is strictly prohibited.

REVISIONS

REV.	ECO	DESCRIPTION	BY	DATE
E	7045	Release Product Data Sheet	Jayne Trost	07/10/2015
F	8008	Update card with formatting changes.	Jayne Trost	2018-08-16


Manufacturer: Laserquick Newberg (or DCI approved equivalent)

Phone No.: 503-537-0824

Website: www.laserquick@earthlink.net

Manufacturer P/N: 91702

Artwork Information: Finished size 4" x 2.5" (no scaling of image), print 2-sided on white card stock, no color images



REPLACEMENT INSTRUCTIONS

Removing Old Parts

1. Remove the following parts from the syringe head:
 - a. DCI Tip Adapter Assembly #3065; using the 5/32 hex wrench provided, insert it in the front of the Quick-Change™ Tip Adapter and turn counterclockwise to remove.
 - b. Tip Adapter Spring #9677.
 - c. Tip Adapter Pin #9659; using needle nose pliers or hemostats, pull straight out to remove.
 - d. Syringe Head O-ring #2201; remove from recessed seat in syringe head.
2. Inspect internal surfaces of syringe head for debris.
3. Clean internal surfaces of syringe head with a cotton swab soaked in alcohol.

Installing New Parts

1. Insert O-ring #2201 in recessed seat in syringe head and lightly lubricate with DCI lubricant #8031.
2. Install Tip Adapter Pin #9659 in hole located in syringe head.
3. Place Tip Adapter Spring #9677 in syringe head making sure it is centered.
4. Locate slot on outside ring of the Quick-Change™ Tip Adapter and insert the adapter in the syringe head, lining up the slot with the Tip Adapter Pin.
5. Using the 5/32 hex wrench, screw in the Tip Adapter Assembly #3065 until firmly in place.
6. Check the operation of the Quick-Change™ Tip Adapter by pushing in the outside ring to make sure it moves freely and holds the syringe tip in place.

All components and assemblies must be RoHS compliant.

Not applicable to this product.

DCI NEWBERG
OREGON

APPROVAL:

J. Trost

DATE:

2018-08-16

TITLE:

Instructions, Replacement Kit, Quick-Change Adapter 3089

PART NO.:

91702

REV.:

F

SHEET

1

OF

1